

Section III: Schedule of Requirements

Technical specifications for Goods and Comparative Data Table:

Lot 1: Eight 08 Airway disinfection apparatus

Item No	UNOPS minimum technical requirements	Quantity	Is quotation compliant? Bidder to complete	Details of goods offered. Bidder to complete
1.1	<p>Mobile and fully automatic aerosol generator intended for the disinfection of surfaces of previously cleaned medical devices. Use with disinfection products.</p> <ul style="list-style-type: none"> ➤ Maximum volume to be treated: 220 - 250 m3 ➤ Average flow rate: 950 - 1000 ml / h; ➤ Sealed membrane compressor; ➤ Helical fan: 230 - 270 m3 / h; ➤ Reserve capacity: 5 liter container; ➤ Empty weight: 25 - 32 kg; ➤ Power supply: 220 V - 50 Hz; ➤ Power: 500 - 700 VA <p>Warranty: minimum 2 years Manual available in french</p>	8	<input type="checkbox"/> Yes <input type="checkbox"/> No	Insert details of goods offered, including specifications and brand/model offered if applicable

Lot 2: Twelve 12 Airway disinfection apparatus

Item No	UNOPS minimum technical requirements	Quantity	Is quotation compliant? Bidder to complete	Details of goods offered. Bidder to complete
2.1	<p>Mobile and fully automatic aerosol generator intended for the disinfection of surfaces of previously cleaned medical devices. Use with disinfection products.</p>	12	<input type="checkbox"/> Yes <input type="checkbox"/> No	Insert details of goods offered, including specifications and brand/model offered if applicable

	<ul style="list-style-type: none"> ➤ Maximum volume to be treated: 220 - 250 m3 ➤ Average flow rate: 950 - 1000 ml / h; ➤ Sealed membrane compressor; ➤ Helical fan: 230 - 270 m3 / h; ➤ Reserve capacity: 5 liter container; ➤ Empty weight: 25 - 32 kg; ➤ Power supply: 220 V - 50 Hz; ➤ Power: 500 - 700 VA. <p>Warranty: minimum 2 years Manual available in french</p>			
--	---	--	--	--

Lot 3: Sixteen 16 Airway disinfection apparatus

Item No	UNOPS minimum technical requirements	Quantity	Is quotation compliant? Bidder to complete	Details of goods offered. Bidder to complete
3.1	<p>Mobile and fully automatic aerosol generator intended for the disinfection of surfaces of previously cleaned medical devices. Use with disinfection products.</p> <ul style="list-style-type: none"> ➤ Maximum volume to be treated: 220 - 250 m3 ➤ Average flow rate: 950 - 1000 ml / h; ➤ Sealed membrane compressor; ➤ Helical fan: 230 - 270 m3 / h; ➤ Reserve capacity: 5 liter container; ➤ Empty weight: 25 - 32 kg; ➤ Power supply: 220 V - 50 Hz; ➤ Power: 500 - 700 VA <p>Warranty: minimum 2 years Manual available in french</p>	16	<input type="checkbox"/> Yes <input type="checkbox"/> No	<p>Insert details of goods offered, including specifications and brand/model offered if applicable</p>

Delivery requirements and Comparative Data Table:

UNOPS Requirements		Is quotation compliant? Bidder to complete	Details Bidder to complete
Delivery time	Bidder shall deliver the goods 6 weeks FCA + 2 weeks DAP after Contract signature.	<input type="checkbox"/> Yes <input type="checkbox"/> No	Insert details
Delivery place and Incoterms rules	Delivery should be made in Bamako, MALI and Incoterms rules as per DAP Incoterms 2020.	<input type="checkbox"/> Yes <input type="checkbox"/> No	Insert details
Consignee details	Organisation Mondiale de la Santé (OMS) au Mali Bureau de Représentation du Mali Quartier N'Tomikorobougou (en face de la Direction Générale de la Santé), BP 99 Bamako, Mali	<input type="checkbox"/> Yes <input type="checkbox"/> No	Insert details
UNOPS Right to vary requirements	At the time the Contract is awarded, UNOPS reserves the right to vary the quantity of the goods and associated services specified above, provided this does not exceed +/- 20%, without any change in the unit prices or other terms and conditions of the RFQ.	<input type="checkbox"/> Yes <input type="checkbox"/> No	Insert details