

DIGITAL PORTABLE X-RAY SYSTEM 2 PIECES

Technical specification

- Power supply: 220B,50/60 Hz
- Ability to be transported between hospital wards
- Have a rechargeable battery
- Output power: 5kW , 110 kW, not less than100 mA
- X-ray tube voltage: 40-110 kW /1kW with step
- X-ray tube current : 25 mA-200 mA
- Console&Display : LCD mAs, κB, mA (to show)
- Principle of operation: Wired / wireless control
- Have wheels and brakes suitable for transporting the equipment
- The power cable should be longer than 5 meters
- Have a digital image recorder
- Have video software

X-ray protection equipments

- X-ray protective apron-1 piece
 - X-ray protective glasses-1 piece
-

ДИЖИТАЛ РЕНТГЕН АППАРАТ /ЗӨӨВРИЙН/-2 ШИРХЭГ

Техникийн үндсэн үзүүлэлт

- Цахилгааны тэжээл: 220В, 50/60Гц байх
- Эмнэлгийн тасаг хооронд зөөвөрлөх боломжтой байх
- Цэнэглэгддэг батарейтай байх
- Гаралтын чадал : 5кВт , 110 кВ, 100 мА -аас багагүй байх
- Хоолойн хүчдэл : 40-110 кВ /1кВ алхамтай/ байх
- Хоолойн гүйдэл : 25 мА-200 мА байх
- Дэлгэц : LCD mAs, кВ, мА харуулдаг байх
- Ажиллах зарчим : Утастай/ утасгүй удирдлага байх
- Төхөөрөмжийг зөөвөрлөхөд тохиромжтой дугуй, тоормозны системтэй байх
- Тэжээлийн кабель 5 метрээс урт байх
- Дижитал дүрс бүртгэгчтэй байх
- Дүрсний програм хангамжтай байх

Рентген хамгаалалтын хэрэгсэл

- Хормогч-1ш
- Нүдний шил-1ш