

**UNHCR**United Nations High Commissioner for Refugees
Haut Commissariat des Nations Unies pour les réfugiés**DATE: 02/11/2021****INVITATION TO BID: No. ITB/HCR/MW/05/2021****FOR THE SUPPLY OF HOSPITAL EQUIPMENT****ANNEX A – TECHNICAL SPECIFICATIONS**

	ITEM	SPECIFICATION	QUANTITY
1	Hospital Beds	<ul style="list-style-type: none">• Multi-function ICU Beds• Height Adjustment• Collapsible side rails• Wheels with brakes for easy movement	43
2	Hospital Beds	<ul style="list-style-type: none">• Strong and steady structure• Single crank with stainless steel heads	20
3	Hospital Mattresses	<ul style="list-style-type: none">• At least 150mm thickness• Sleeping surface of approximately 900 x 2000 mm• enclosed by a removable waterproof cover	63
4	Pillows	<ul style="list-style-type: none">• Made of antibacterial. Anti- static, toxic free, high density foam core• Removable cotton protection cover branded with UNHCR Logo	63
5	Drip Stands	<ul style="list-style-type: none">• Stable and strong stainless-steel structure• Height adjustable (approx. 140 -200cm)• Large base of 5 feet steel (approx. 60cm diameter)• Swivel castors. 2 with Brakes• Complete with quadruple hook for infusion bags	63
6	Maternity Gynecology Labor Delivery Mattresses	<ul style="list-style-type: none">• At least 150mm thickness• Sleeping surface of approximately 900 x 2000 mm• enclosed by a removable waterproof cover	4
7	Medical Instrument Trolleys	<ul style="list-style-type: none">• Stainless steel• Three shelves• Mounted on 4 castors	3
8	Medical Dressing Trolley	<ul style="list-style-type: none">• Stainless steel• Two shelves• Mounted on 4 castors	1
9	Laundry Trolleys (on Castors)	<ul style="list-style-type: none">• Trolley made of stainless steel.• 3 S.S. shelves with guard rails.• Supplied with 1 bag for dirty linen.• 100mm Swivel castor, 2 with brake	2
10	Drug Trolley	<ul style="list-style-type: none">• Framework made of stainless steel• Stainless steel rails covering sides of the shelves• 2 shelves• mounted on four swivel castors (2 with brakes)	1
11	Patient Trolley	<ul style="list-style-type: none">• Aluminum alloy material.	1

		<ul style="list-style-type: none"> • Four universal swivel wheels. • Bed surface of the stretcher is made of foamed cushion, waterproof. • adjustable back rest. • With straps for binding patients • Max. Angle of the Back:70° 	
12	Examination Couch	<ul style="list-style-type: none"> • Maximal load: 150kg. • Frame: Stainless steel tubes, Ø 30mm. • Upholstery: simulated leather, X-ray transparent. • Mattress, fixed, without seams, 40mm thick. • Headrest: adjustable automatic locking system. • Equipped with paper roll holder 	1
13	Stand Bowl (kickabout on Castors)	<ul style="list-style-type: none"> • Double Bowl Holding • Stainless Steel tubular pipe construction mounted on heavy polypropylene base with castors. • Supplied complete with 2 Stainless Steel basins of 35 cm. diameter. 	2
14	Surgeon's Stool (Adjustable Height)	<ul style="list-style-type: none"> • Chrome base • Adjustable height swivel • Cushion diameter 35cm • Antistatic imitation leather 	3
15	Instrument Cabinet	<ul style="list-style-type: none"> • Stainless Steel Construction • Full-Height Storage Cabinet • Shatter-Proof Safety Glass in Doors • Full-View Doors for Complete Visual Inspection • Adjustable Shelves • Locking Latch 	1
16	Drugs Cupboard (Lockable)	<ul style="list-style-type: none"> • Medicine storage cabinet, double door • Mounted on 4 sturdy supports, all finished with height adjustable feet • Plain side panels and doors, block view on cabinet content • Inside reinforced fixtures facilitate height adjustment of the 4 shelves • Doors are triple hinged, and lockable with key • All doors have recessed finger grip-hold along their entirely length, leaving front flat • Clearance underneath the cabinet allows for easy cleaning 	1
17	Dangerous Drugs Cupboard (Lockable)	<ul style="list-style-type: none"> • Height adjustable internal shelves. • Single point locking, one lock supplied with three keys. 	1
18	Drug Box (Lockable)	<ul style="list-style-type: none"> • With different storage units to classify and store medicines. • Durable use, not easily to break in. • Double safety lock 	1
19	Drug Refrigerator	<ul style="list-style-type: none"> • Lockable • Audible and visual high and low temperature alarms. • Open door alarm • Front lockable roller castors. • Stainless steel shelves 	2

		<ul style="list-style-type: none"> • Temperature Monitoring Software. • Power Failure Alarm 	
20	Examination Lights	<ul style="list-style-type: none"> • Adjustable • Angle Poise • LED 	4
21	Bed Sheets	<ul style="list-style-type: none"> • Cotton Material • 150cm x 240 cm • Branded with UNHCR Logo 	500
22	Bed Covers	<ul style="list-style-type: none"> • 100% Cotton • Woven type • 195cm x 220cm • Branded with UNHCR Logo 	500
23	Operating Table	<ul style="list-style-type: none"> • Side controlled • Divided leg plate • Ministry prize • Stainless steel bottom cover 	1
24	Theatre Lights	<ul style="list-style-type: none"> • Light Source Type: Halogen. • Number Of Domes: 3, 2. • Movement: Angular. • Life Span Of Leds: 50,000 hrs. • Color Temperature: 4300 K. • Focus Diameter: 15-25 cm. • Light Intensity: 150,000 Lux. 	1
25	Kangaroo Care Room Heater	<ul style="list-style-type: none"> • Articulable irradiant reflector with observation light • four wheels with brake • A very stable system with complete all-round access during critical situation • Barriers for infant protection • bed tilt of 12 degrees on either side • IV pole and infusion pump can be clamped • central utility tray 	16

ANNEX A – TECHNICAL EVALUATION CRITERIA

	Detailed criteria	Scoring method
1	Valid company registration certificates (certificate of incorporation) and valid tax compliance certificate.	Pass or Fail
2	Proof of experience in the supply and delivery of similar items. (Copies of Completed Purchase Orders and/or References from reputable institutions)	Pass or Fail
3	A catalogue of items being offered. State source of the goods and compliance with the proposed technical specifications.	Pass or Fail
4	An outline/chart of proposed delivery timelines	Pass or Fail