

Section II: Schedule of Requirements

Summary of Requirements


UNOPS requirements are comprised of the following Lots:

- Lot 1: Bedside Locker - 346 EA
- Lot 2: Examination table - 12 EA
- Lot 3: ICU Bed (Electric) - 268 EA
- Lot 4: ICU Bed (Manual) - 78 EA
- Lot 5: IV Stand - 740 EA
- Lot 6: Cardiac table - 346 EA
- Lot 7: Wheelchair - 1348 EA

Lot No 1: Bedside Locker - Quantity: 346 EA

Technical specifications for Goods:

No	UNOPS minimum technical requirements
1	<u>Bedside locker</u> Bedside cabinets for hospital ward.
1.1	<u>Operational Characteristics</u>
1.1.1	Body and structure made of Stainless steel, at least 14 gauge or laminated melamine/particle board with aluminium frame.
1.1.2	The hospital bed cabinet must have: <ul style="list-style-type: none"> • Top surface. • Cupboard section. • Pull out the drawer for ease of cleaning. • Swing door with locking/latching device.
1.1.3	Swing door and drawer shall have suitable handles.
1.1.4	With 4 dual casters at least 60 mm diameter (at least two with brakes).
1.1.5	With electrostatic powder or epoxy coating in case of stainless steel cabinets and colour laminate in case of other material. Vendor to submit the colour choices available from the manufacturer.
1.1.6	Top surface should be with raised edges or perimeter rails for safe storage of items.
1.1.7	No toxic construction materials.
1.2	<u>Approximate dimensions</u>
1.2.1	The cupboard section shall be around 40 cm high.
1.2.2	45 × 47 × 90 cm (W×D×H) +/-5%
1.3	<u>Support Services</u>
1.3.1	Manufacturer's warranty: <ol style="list-style-type: none"> a) Duration 24 months from final receipt covering all manufacturing defects. b) The warranty covers all delivered equipment and the accessories delivered.
1.3.2	Warranty with scope extended by the bidder: <ol style="list-style-type: none"> a. Duration 24 months from final receipt covering all manufacturing defects. b. It includes corrective maintenance: labour and spare parts with no call limitations with the sole exclusion of damage due to proven misuse. c. Corrective maintenance during warranty can be carried out at the beneficiary premise, or by mailing the broken equipment at supplier or manufacturer premise and mailing back the repaired equipment or the substituting one to the Beneficiary premise. In the case of transportation all the costs and risks will be covered by the Supplier. The time elapsed between the communication of the faulty equipment and the intervention on site will be, within the warranty period, no more than 2 business days.


	d. The Supplier will notify the end user in case of any equipment Recall / Field safety Notice / Alerts as issued by the manufacturer/NRA or any other relevant authorities and organisations
1.4	<u>Documents to be presented with the offer</u>
1.4.1	Bidder declaration of the brand name, manufacturer, Country of Origin and model name of all the goods (including all consumables and accessories included in the offer).
1.4.2	Manufacturer's ISO 13485 certificate or ISO 9001 when ISO13485 is not applicable.
1.4.3	Manuals: Copies of User manual must be submitted in digital format in English.
1.4.4	Manufacturer's technical marketing catalogues. Manufacturer marketing catalogues of the goods offered to verify compliance with the technical specifications. Catalogues must be provided in English.
1.4.5	Bidder shall provide invoice/proforma invoice, packing list and at least a draft shipping document without indicating the word "Draft" and with fixed BL/AWB number to UNOPS at least one month in advance before the actual shipping date for the application of Tax Exemption Certificate (TEC).
1.5	<u>Payment Terms:</u>
1.5.1	<ul style="list-style-type: none"> 80% of the cost of the goods shall be paid upon delivery per the Incoterms 2020 20% of the cost of the goods and 100% of the cost of associated services shall be paid upon final acceptance of the goods and the associated services (delivery, installation, testing, commissioning, training of users on operation), as applicable.
1.5.2	Sample Photo - The image below is intended solely as a guiding support and should be considered as purely indicative and not restrictive of the expected item characteristics:
	

Delivery requirements for bedside locker:

Delivery time	Bidder shall deliver the goods within 10 weeks after issuance of Purchase Order to UNOPS warehouse in Yangon, Myanmar (excluding the time for custom clearance process).
Delivery place and Incoterms rules	UNOPS warehouse in Yangon, Myanmar. Delivered at Place Unloaded [DPU] Incoterms 2020
Consignee details	UNOPS Myanmar No.12/O , Pyi Thu Lane, 7 Miles, Mayangone T/S, Yangon, Myanmar
UNOPS Right to vary requirements	At the time the Contract is awarded, UNOPS reserves the right to vary the quantity of the goods and associated services specified above, provided this does not exceed +/- 20%, without any change in the unit prices or other terms and conditions of the ITB.
Pre-inspection	The goods will be pre-inspected by UNOPS or by a third party appointed by UNOPS either in the production site or in Yangon.

Lot No 2: Examination table - Quantity: 12 EA
Technical specifications for Goods:

No	UNOPS minimum technical requirements
2	<u>Examination table</u> General examination table suitable for physician office and minor procedures.
2.1	<u>Operational Characteristics</u>
2.1.1	Strong sturdy steel frame structure at least AISI 305 steel.
2.1.2	Fixed height.
2.1.3	Two sections.
2.1.4	Adjustable back section by lever: 0 to 60 degrees.
2.1.5	Epoxy coated frame structure.
2.1.6	Padded with expandable polyurethane foam rubber, with a density of at least 24 kg/m3 and at least 3 mm thick.
2.1.7	Seamless upholstery antibacterial, antistatic, flame-retardant, resistant to corrosion of water, detergent soap, ethyl alcohol solution and to the hypochlorite of sodium.
2.1.8	Couch roll holder bracket.
2.1.9	Fixed on strong stable legs with anti static rubber feet.
2.1.10	Dimensions (approximate): 190 x 55 x 80 cm (LxWxH).
2.1.11	Load capacity not below 130 kg.
2.2	<u>Accessories to be included in the offered price</u>
2.2.1	Two (2) rolls of paper.
2.3	<u>Support Services</u>
2.3.1	Equipment installation: the equipment must be delivered at destination, already assembled and tested.
2.3.2	Training on use and maintenance: Operating, cleaning, daily maintenance and training instructions should be included in the manuals delivered with the equipment.
2.3.3	Manufacturer's warranty: a) Duration 24 months from final receipt covering all manufacturing defects. b) The warranty covers all delivered equipment and the accessories delivered.
2.3.4	Warranty with scope extended by the bidder: a) Duration 24 months from final receipt covering all manufacturing defects. b) It includes corrective maintenance: labour and spare parts with no call limitations with the sole exclusion of damage due to proven misuse. c) The preventive and corrective maintenance technical service must be performed by a company based in the country of destination. d) The Supplier will notify the end user in case of any equipment Recall / Field safety Notice / Alerts as issued by the manufacturer/NRA or any other relevant authorities and organizations
2.4	<u>Documents to be presented with the offer</u>
2.4.1	Bidder declaration of the brand name, manufacturer, Country of Origin and model name of all the goods (including all consumables and accessories included in the offer)
2.4.2	Manufacturer's ISO 13485 certificate or ISO 9001 when ISO13485 is not applicable.
2.4.3	Market authorization for one of the following markets/requested certifications: a) EUROPE/CE 2017/745 conformity or equipment with CE mark according with 93/42 EEC b) USA/FDA certification (Food and Drug Administration) for internal US market c) CANADA/SOR/98-282 conformity certification d) AUSTRALIA/TGA Conformity certification

	e) JAPAN/ PMDA pre-market approval
2.4.4	Certificate of registration/Authorization/ Waiver the Import and use of the device by the health regulatory authority (NRA) of the Country of destination.
2.4.5	Factory certificate that the equipment, and specifically the offered model, is not discontinued on the production line.
2.4.6	Manufacturer's Authorization: Bidders may submit a simple copy of the manufacturer's authorization within their bid, indicating the offered model, according to the bidder's own form.
2.4.7	Manufacturer's technical marketing catalogues. Manufacturer marketing catalogues of the goods offered to verify compliance with the technical specifications. Catalogues must be provided in English.
2.4.8	Bidder shall provide invoice/proforma invoice, packing list and at least a draft shipping document without indicating the word "Draft" and with fixed BL/AWB number to UNOPS at least one month in advance before the actual shipping date for the application of Tax Exemption Certificate (TEC).
2.5	<u>Payment Terms</u>
2.5.1	<ul style="list-style-type: none"> 80% of the cost of the goods shall be paid upon delivery per the Incoterms 2020 20% of the cost of the goods and 100% of the cost of associated services shall be paid upon final acceptance of the goods and the associated services (delivery, installation, testing, commissioning, training of users on operation), as applicable.
2.5.2	Sample Photos - The images below are intended solely as a guiding support and should be considered as purely indicative and not restrictive of the expected item characteristics:
	

Delivery requirements:

Delivery time	Bidder shall deliver the goods within 10 weeks after issuance of Purchase Order to UNOPS warehouse in Yangon, Myanmar (excluding the time for custom clearance process).
Delivery place and Incoterms rules	UNOPS warehouse in Yangon, Myanmar. Delivered at Place Unloaded [DPU] Incoterms 2020
Consignee details	UNOPS Myanmar No.12/O , Pyi Thu Lane, 7 Miles, Mayangone T/S, Yangon, Myanmar
UNOPS Right to vary requirements	At the time the Contract is awarded, UNOPS reserves the right to vary the quantity of the goods and associated services specified above, provided this does not exceed +/- 20%, without any change in the unit prices or other terms and conditions of the ITB.
Pre-inspection	The goods will be pre-inspected by UNOPS or by a third party appointed by UNOPS either in the production site or in Yangon.

Related services requirements for Examination Table:

SN	UNOPS minimum requirements for services	Place where services will be performed	Quantity	Physical unit
1	Assembled at destinations	Ann District Hospital	3	EA
2	Assembled at destinations	Moe Nyin District Hospital	3	EA

SN	UNOPS minimum requirements for services	Place where services will be performed	Quantity	Physical unit
3	Assembled at destinations	Kutkai General Hospital -Shan North	3	EA
4	Assembled at destinations	Shadaw General Hospital - Kayah State	3	EA
Total quantity			12	

Lot No 3: ICU Bed (Electric) - Quantity: 268 EA
Technical specifications for Goods:

No	UNOPS minimum technical requirements
3	<u>ICU Bed (Electric)</u> Electrical dual-sided mattress platform, height adjustable, 3 articulated sections, bed for intensive care patients.
3.1	<u>Technical requirements:</u>
3.1.1	Electrical requirements: 220V±10% 50 Hz mono-phase electrical source, supplied with hospital grade power BS 1363 plug.
3.1.2	Frame:
3.1.2 (a)	Made of 4 epoxy-coated steel at least 12 gauge, round tube legs including seats for wheels and locking gears.
3.1.2 (b)	Equipped with plastic bumpers and receptacles for patient-lifting and/or IV poles.
3.1.2 (d)	Safe working load not less than 130 Kg.
3.1.2 (e)	Overall dimension: 2±0.1 × 0.9±0.1 m (L x W).
3.1.2 (f)	Two (2) folding or tilting safe stainless steel side rails with locking mechanism, at least 3/4 bed size.
3.1.2 (g)	All construction materials should be highly resistant to corrosion like stainless steel, epoxy coated steel or similar.
3.1.3	Mattress platform:
3.1.3 (a)	Base consisting of rigid non conductive hard plastic surfaces to allow save cardiopulmonary resuscitation.
3.1.3 (b)	Back section must be made of radiolucent material and shall include an x-ray cassette holder.
3.1.3 (c)	Mattress retainers.
3.1.3 (d)	Radiotransparent.
3.1.4	Mattress and pillow:
3.1.4 (a)	Fire resistant mattress, made of high-density polyurethane foam (at least 24 kg / m3) with waterproof coating, made of hypoallergenic material.
3.1.4 (b)	Mattress size according to the size of the bed.
3.1.4 (c)	Removable waterproof, antibacterial, anti-static, toxic-free and seamless cover.
3.1.4 (d)	Mattress thickness of at least 120 mm.
3.1.4 (e)	Pillow made of antibacterial, anti-static, toxic-free, high-density foam core, provided with removable and washable hard cotton cover.
3.1.4 (f)	Mattress cover and pillow resistant to hospital cleaning agents: water and detergent soap; 70% ethyl alcohol solution with or without nitrite; sodium hypochlorite; hospital use solvents and biological liquids.
3.1.5	Movements/controls: a) Allowed movement / positions: Fowler / Semi-fowler, Trendelenburg / Anti-trendelenburg and knee lift, raising and keeping the segment horizontal from knee to ankle. b) Backrest movement adjustment range of at least: 0 to 65° c) Knee elevation of at least: 0 to 30° d) Trendelenburg / Anti-trendelenburg of at least: + 12°/-12°.

No	UNOPS minimum technical requirements
	e) Cardiopulmonary resuscitation (CPR) position required, with single button or manual control. f) Vertical movement control/adjustment via mechanically operated actuator with dual-sided pedal control. g) Vertical movement range of at least between 60 and 75 cm. h) Easily removable head & foot ends with plastic-covered bends and grips.
3.2	<u>Standards:</u>
3.2.1	Manufacturer declaration of painting thickness and resistance conforming to UNI-EN-ISO 2409 and ISO-2802.
3.2.2	WELDING: Manufacturer declaration of conformity to UNI-EN 288 and UNI-EN 25817.
3.3	<u>Accessories to be included in the offered price:</u>
3.3.1	Two (2) stainless steel telescopic IV holders with 2 hooks each.
3.3.2	Support for oxygen tubes.
3.3.3	Possibility of placing trauma accessories.
3.3.4	Blanket support.
3.3.5	Chart holder.
3.3.6	Urine/Liquids sack holder.
3.4	<u>Environmental Conditions:</u>
3.4.1	Operating conditions range: at least 5°C - 35°C / 20% – 90% RH.
3.4.2	Storage conditions range: at least -10°C - 55°C / 20 % – 80% RH.
3.4.3	Atmospheric pressure operating range: at least 800 ~ 1060hPa.
3.4.4	Protection rating: at least IP55 or better.
3.5	<u>Support Services</u>
3.5.1	Equipment installation: the equipment must be assembled and tested (functional and electrical safety test) at the destination unit and the results, certification signed by a qualified technician, of the functional and electrical safety tests delivered to the beneficiary unit.
3.5.2	Training on use and maintenance: a) User and Maintenance manuals delivered with each equipment. b) Training on site for users and maintenance people. c) User training shall include at least: 1) Normal operating mode, 2) Daily and periodic cleaning and maintenance, 3) Basic troubleshooting, 4) Min risks in the use of the equipment. d) Maintenance training shall include at least: 1) Normal operating mode, 2) Daily and periodic cleaning and maintenance, 3) Advanced troubleshooting, 4) Accessories and spare parts description and codes, 5) Warranty conditions, 6) Supplier and Manufacturer contact references. e) The bidder must certify that the technician who will provide the virtual training to the medical and technical personnel was trained by the manufacturer of the equipment. f) The technician must prepare a record of the persons trained with the signatures to confirm the training received.
3.5.3	Manufacturer's warranty: The warranty covers all delivered equipment and accessories. 24 months from final receipt covering all manufacturing defects.
3.5.4	Warranty with scope extended by the bidder: a) The warranty covers all equipment and accessories delivered. b) Duration 24 months including preventive and corrective maintenance at the installation site.

No	UNOPS minimum technical requirements
	<p>c) It includes corrective maintenance: labour and spare parts with no call limitations with the sole exclusion of damage from misuse.</p> <p>d) The time elapsed between the communication of the faulty equipment and the intervention on site will be, within the warranty period, no more than 2 business days.</p> <p>e) The corrective maintenance technical service must be performed by a company based in country of destination</p> <p>f) Notify the end user in case of any equipment Recall / Field safety Notice / Alerts as issued by the manufacturer/NRA or any other relevant authorities and organizations</p>
3.6	<u>Documents to be presented with the offer</u>
3.6.1	Bidder declaration of the brand name, manufacturer, Country of Origin and model name of all the goods (including all consumables and accessories included in the offer)
3.6.2	Manufacturer's ISO 13485 certificate or ISO 9001 when ISO13485 is not applicable.
3.6.3	Market authorization for one of the following markets/requested certifications: <ul style="list-style-type: none"> a) EUROPE/CE 2017/745 conformity or equipment with CE mark according with 93/42 EEC b) USA/FDA certification (Food and Drug Administration) for internal US market c) CANADA/SOR/98-282 conformity certification d) AUSTRALIA/TGA Conformity certification e) JAPAN/ PMDA pre-market approval
3.6.4	Certificate of registration/Authorization/ Waiver the Import and use of the device by the health regulatory authority (NRA) of the Country of destination.
3.6.5	Manufacturer's Authorization: Bidders may submit a simple copy of the manufacturer's authorization within their bid, indicating the offered model, according to the bidder's own form.
3.6.6	Manuals: Copies of User and Maintenance manual must be submitted in digital format in English.
3.6.7	Manufacturer's technical marketing catalogues. Manufacturer marketing catalogues of the goods offered to verify compliance with the technical specifications. Catalogues must be provided in English.
3.6.8	Letter of the manufacturer stating his commitment of spare parts availability for 7 years in case of award (only availability the supply of spare parts shall not be included in the bid price).
3.6.9	Bidder shall provide invoice/proforma invoice, packing list and at least a draft shipping document without indicating the word "Draft" and with fixed BL/AWB number to UNOPS at least one month in advance before the actual shipping date for the application of Tax Exemption Certificate (TEC).
3.7	<u>Payment Terms</u>
3.7.1	<ul style="list-style-type: none"> • 80% of the cost of the goods shall be paid upon delivery per the Incoterms 2020 • 20% of the cost of the goods and 100% of the cost of associated services shall be paid upon final acceptance of the goods and the associated services (delivery, installation, testing, commissioning, training of users on operation), as applicable.

Delivery requirements for ICU Bed (Electric):

Delivery time	Bidder shall deliver the goods within 10 weeks after issuance of Purchase Order to UNOPS warehouse in Yangon, Myanmar (excluding the time for custom clearance process).
Delivery place and Incoterms rules	UNOPS warehouse in Yangon, Myanmar. Delivered at Place Unloaded [DPU] Incoterms 2020
Consignee details	UNOPS Myanmar No.12/O , Pyi Thu Lane, 7 Miles, Mayangone T/S, Yangon, Myanmar
UNOPS Right to vary requirements	At the time the Contract is awarded, UNOPS reserves the right to vary the quantity of the goods and associated services specified above, provided this does not exceed +/- 20%, without any change in the unit prices or other terms and conditions of the ITB.

Pre-inspection	The goods will be pre-inspected by UNOPS or by a third party appointed by UNOPS either in the production site or in Yangon.
-----------------------	---

Related services requirements for ICU Bed (Electric):

SN	UNOPS minimum requirements for services	Place where services will be performed	Qty	Physical unit
1	Assembled at destination and on site training for user and maintenance people.	Yangon General Hospital	20	EA
2	Assembled at destination and on site training for user and maintenance people.	Mandalay General Hospital	16	EA
3	Assembled at destination and on site training for user and maintenance people.	North Okkala General Hospital	12	EA
4	Assembled at destination and on site training for user and maintenance people.	Naypyitaw General Hospital	10	EA
5	Assembled at destination and on site training for user and maintenance people.	Yangon Specialist Hospital (500-bed)	10	EA
6	Assembled at destination and on site training for user and maintenance people.	Thingangyun General Hospital	10	EA
7	Assembled at destination and on site training for user and maintenance people.	Magway General Hospital, Magway Region	10	EA
8	Assembled at destination and on site training for user and maintenance people.	Insein General Hospital	8	EA
9	Assembled at destination and on site training for user and maintenance people.	Magway Teaching Hospital	6	EA
10	Assembled at destination and on site training for user and maintenance people.	Sao San Tun General Hospital	8	EA
11	Assembled at destination and on site training for user and maintenance people.	Women and Children Hospital, Shan South State	4	EA
12	Assembled at destination and on site training for user and maintenance people.	300-bed Mandalay Teaching Hospital	10	EA
13	Assembled at destination and on site training for user and maintenance people.	Myitkyina General Hospital, Kachin State	6	EA
14	Assembled at destination and on site training for user and maintenance people.	Bamaw General Hospital, Kachin State	4	EA
15	Assembled at destination and on site training for user and maintenance people.	Loikaw General Hospital, Kayah State	6	EA
16	Assembled at destination and on site training for user and maintenance people.	Hpa-an General Hospital, Kayin State	6	EA
17	Assembled at destination and on site training for user and maintenance people.	Hakha General Hospital, Chin State	4	EA
18	Assembled at destination and on site training for user and maintenance people.	Monywa General Hospital, Sagaing Region	6	EA
19	Assembled at destination and on site	Kalay General Hospital, Sagaing Region	6	EA

SN	UNOPS minimum requirements for services	Place where services will be performed	Qty	Physical unit
	training for user and maintenance people.			
20	Assembled at destination and on site training for user and maintenance people.	Sagaing General Hospital, Sagaing Region	4	EA
21	Assembled at destination and on site training for user and maintenance people.	Dawei General Hospital, Tanintharyi Region	6	EA
22	Assembled at destination and on site training for user and maintenance people.	Myeik General Hospital, Tanintharyi Region	6	EA
23	Assembled at destination and on site training for user and maintenance people.	Bago General Hospital, Bago Region	6	EA
24	Assembled at destination and on site training for user and maintenance people.	Pyay General Hospital, Bago Region	6	EA
25	Assembled at destination and on site training for user and maintenance people.	Taunggu General Hospital, Bago Region	4	EA
26	Assembled at destination and on site training for user and maintenance people.	Minbu General Hospital, Magway Region	4	EA
27	Assembled at destination and on site training for user and maintenance people.	Pakkoku General Hospital, Magway Region	4	EA
28	Assembled at destination and on site training for user and maintenance people.	Pyinoolwin General Hospital, Mandalay Region	4	EA
29	Assembled at destination and on site training for user and maintenance people.	Meikthila General Hospital, Mandalay Region	4	EA
30	Assembled at destination and on site training for user and maintenance people.	Mawlamyine General Hospital, Mon State	6	EA
31	Assembled at destination and on site training for user and maintenance people.	Mawlamyine Women and Children Hospital, Mon State	4	EA
32	Assembled at destination and on site training for user and maintenance people.	Sittwe General Hospital, Rakhine State	6	EA
33	Assembled at destination and on site training for user and maintenance people.	East Yangon General Hospital, Yangon Region	6	EA
34	Assembled at destination and on site training for user and maintenance people.	West Yangon General Hospital, Yangon Region	6	EA
35	Assembled at destination and on site training for user and maintenance people.	Hlaingtharyar General Hospital, Yangon Region	6	EA
36	Assembled at destination and on site training for user and maintenance people.	Thanlyin General Hospital, Yangon Region	6	EA
37	Assembled at destination and on site training for user and maintenance people.	Lashio General Hospital, Shan North State	6	EA
38	Assembled at destination and on site training for user and maintenance people.	Kengtong General Hospital, Shan East State	6	EA

SN	UNOPS minimum requirements for services	Place where services will be performed	Qty	Physical unit
39	Assembled at destination and on site training for user and maintenance people.	Patheingyi General Hospital, Ayeyarwady Region	6	EA
Total quantity			268	EA

Lot No 4: ICU Bed (Manual) - Quantity: 78 EA
Technical specifications for Goods:

No	UNOPS minimum technical requirements
4	ICU Bed Hydraulic Hydraulic dual-sided mattress platform, height adjustable, 3 articulated sections, bed for intensive care patients.
4.1	Technical requirements:
4.1.1	Frame:
4.1.1 (a)	Made of 4 epoxy-coated steel at least 12 gauge, round tube legs including seats for wheels and locking gears.
4.1.1 (b)	Equipped with plastic bumpers and receptacles for patient-lifting and/or IV poles.
4.1.1 (c)	All lifting mechanisms and hydraulic pumps support are made of epoxy coated, cold-drawn steel.
4.1.1 (d)	Safe working load not less than 130 Kg.
4.1.1 (e)	Overall dimension: 2±0.1 × 0.9±0.1 m (L x W).
4.1.1 (f)	Two (2) folding or tilting safe stainless steel side rails with locking mechanism, at least 3/4 bed size.
4.1.1 (g)	All construction materials should be highly resistant to corrosion like stainless steel, epoxy coated steel or similar.
4.1.2	Mattress platform:
4.1.2 (a)	Base consisting of rigid non conductive hard plastic surfaces to allow save cardiopulmonary resuscitation.
4.1.2 (b)	Back section must be made of radiolucent material and shall include an x-ray cassette holder.
4.1.2 (c)	Mattress retainers.
4.1.2 (d)	Radiotransparent.
4.1.3	Mattress and pillow:
4.1.3 (a)	Fire resistant mattress, made of high-density polyurethane foam (at least 24 kg / m ³) with waterproof coating, made of hypoallergenic material.
4.1.3 (b)	Mattress size according to the size of the bed.
4.1.3 (c)	Removable waterproof, antibacterial, anti-static, toxic-free and seamless cover.
4.1.3 (d)	Mattress thickness of at least 120 mm.
4.1.3 (e)	Pillow made of antibacterial, anti-static, toxic-free, high-density foam core, provided with removable and washable hard cotton cover.
4.1.3 (f)	Mattress cover and pillow resistant to hospital cleaning agents: water and detergent soap; 70% ethyl alcohol solution with or without nitrite; sodium hypochlorite; hospital use solvents and biological liquids.
4.1.4	Movements/controls:
a)	Allowed movement / positions: Fowler / Semi-fowler, Trendelenburg / Anti-trendelenburg and knee lift, raising and keeping the segment horizontal from knee to ankle.

No	UNOPS minimum technical requirements
	b) Backrest movement adjustment range of at least: 0 to 65°. c) Knee elevation of at least: 0 to 30°. d) Trendelenburg / Anti-trendelenburg of at least: + 12°/-12°. e) Cardiopulmonary resuscitation (CPR) position required, with manual control. f) Vertical movement control/adjustment via mechanically operated actuator with dual-sided pedal control or by crank mechanism g) Vertical movement range of at least between 60 and 75 cm. h) Easily removable head & foot ends with plastic-covered bends and grips.
4.2	<u>Accessories to be included in the offered price:</u>
4.2.1	Two (2) stainless steel telescopic IV holders with 2 hooks each.
4.2.2	Support for oxygen tubes.
4.2.3	Possibility of placing trauma accessories.
4.2.4	Blanket support.
4.2.5	Chart holder.
4.2.6	Urine/Liquids sack holder.
4.3	<u>Support Services:</u>
4.3.1	Equipment installation: the equipment must be delivered at destination, already assembled and tested.
4.3.2	Manufacturer's warranty: The warranty covers all delivered equipment and accessories. Duration 24 months from final receipt covering all manufacturing defects.
4.3.3	Warranty with scope extended by the bidder: a) Duration 24 months from final receipt covering all manufacturing defects. b) It includes corrective maintenance: labour and spare parts with no call limitations with the sole exclusion of damage due to proven misuse. c) The corrective maintenance technical service must be performed by a company based in the country of destination. d) The Supplier will notify the end user in case of any equipment Recall / Field safety Notice / Alerts as issued by the manufacturer/NRA or any other relevant authorities and organizations.
4.4	<u>Documents to be presented with the offer:</u>
4.4.1	Bidder declaration of the brand name, manufacturer, Country of Origin and model name of all the goods (including all consumables and accessories included in the offer).
4.4.2	Manufacturer's ISO 13485 certificate or ISO 9001 when ISO13485 is not applicable.
4.4.3	Market authorization for one of the following markets/requested certifications: a) EUROPE/CE 2017/745 conformity or equipment with CE mark according with 93/42 EEC b) USA/FDA certification (Food and Drug Administration) for internal US market c) CANADA/SOR/98-282 conformity certification d) AUSTRALIA/TGA Conformity certification e) JAPAN/ PMDA pre-market approval
4.4.4	Certificate of registration/Authorization/ Waiver the Import and use of the device by the health regulatory authority (NRA) of the Country of destination.
4.4.5	Manufacturer's Authorization: Bidders may submit a simple copy of the manufacturer's authorization within their bid, indicating the offered model, according to the bidder's own form.

No	UNOPS minimum technical requirements
4.4.6	Manufacturer's technical marketing catalogues. Manufacturer marketing catalogues of the goods offered to verify compliance with the technical specifications. Catalogues must be provided in English.
4.4.7	Letter of the manufacturer stating his commitment of spare parts availability for 7 years in case of award (only availability the supply of spare parts shall not be included in the bid price).
4.4.8	Bidder shall provide invoice/proforma invoice, packing list and at least a draft shipping document without indicating the word "Draft" and with fixed BL/AWB number to UNOPS at least one month in advance before the actual shipping date for the application of Tax Exemption Certificate (TEC).
4.5	<u>Payment Terms</u>
4.5.1	<ul style="list-style-type: none"> 80% of the cost of the goods shall be paid upon delivery per the Incoterms 2020 20% of the cost of the goods and 100% of the cost of associated services shall be paid upon final acceptance of the goods and the associated services (delivery, installation, testing, commissioning, training of users on operation), as applicable.

Delivery requirements for ICU Bed (Manual):

Delivery time	Bidder shall deliver the goods within 10 weeks after issuance of Purchase Order to UNOPS warehouse in Yangon, Myanmar (excluding the time for custom clearance process).
Delivery place and Incoterms rules	UNOPS warehouse in Yangon, Myanmar. Delivered at Place Unloaded [DPU] Incoterms 2020
Consignee details	UNOPS Myanmar No.12/O , Pyi Thu Lane, 7 Miles, Mayangone T/S, Yangon, Myanmar
UNOPS Right to vary requirements	At the time the Contract is awarded, UNOPS reserves the right to vary the quantity of the goods and associated services specified above, provided this does not exceed +/- 20%, without any change in the unit prices or other terms and conditions of the ITB.
Pre-inspection	The goods will be pre-inspected by UNOPS or by a third party appointed by UNOPS either in the production site or in Yangon.

Related services requirements for ICU Bed (Manual):


SN	UNOPS minimum requirements for services	Place where services will be performed	Q-ty	Physical unit
1	Assembled at destination	Pha-Lan General Hospital, Chin State	4	EA
2	Assembled at destination	Shwe Bo General Hospital, Sagaing Region	4	EA
3	Assembled at destination	Kyaukse General Hospital, Mandalay Region	4	EA
4	Assembled at destination	Nyaung-U General Hospital, Mandalay Region	4	EA
5	Assembled at destination	Kyaukphyu General Hospital, Rakhine State	4	EA
6	Assembled at destination	Loilem General Hospital, Shan South State	4	EA
7	Assembled at destination	Maubin General Hospital, Ayeyarwady Region	4	EA
8	Assembled at destination	Myaungmya General Hospital, Ayeyarwady Region	4	EA
9	Assembled at destination	Laputta General Hospital, Ayeyarwady Region	4	EA
10	Assembled at destination	Pyapon General Hospital, Ayeyarwady Region	4	EA
11	Assembled at destination	Hinthada General Hospital, Ayeyarwady Region	4	EA
12	Assembled at destination	Pyinmana General Hospital, Naypyitaw Council	4	EA

SN	UNOPS minimum requirements for services	Place where services will be performed	Q-ty	Physical unit
13	Assembled at destination	Naypyitaw General Hospital (300-bed)	4	EA
14	Assembled at destination	Tedim Township Hospital	2	EA
15	Assembled at destination	Mindat District Hospital	2	EA
16	Assembled at destination	Mogkok Township Hospital	2	EA
17	Assembled at destination	Yamethin District Hospital	2	EA
18	Assembled at destination	Kawthaung District Hospital	2	EA
19	Assembled at destination	Ye Township Hospital	2	EA
20	Assembled at destination	Thaton District Hospital	2	EA
21	Assembled at destination	Thandwe District Hospital	2	EA
22	Assembled at destination	Tachileik District Hospital	2	EA
23	Assembled at destination	Ann District Hospital	2	EA
24	Assembled at destination	Moe Nyin District Hospital	2	EA
25	Assembled at destination	Kutkai General Hospital -Shan North	2	EA
26	Assembled at destination	Shadaw General Hospital - Kayah State	2	EA
Total quantity			78	EA

Lot No 5: IV Stand - Quantity: 740 EA
Technical specifications for Goods:

No	UNOPS minimum technical requirements
5	<u>IV Stand</u> Mobile Support for intravenous fluid delivery.
5.1	<u>Technical requirements</u>
5.1.1	Manufactured from 12 gauge polished stainless steel.
5.1.2	Four (4) infusion bag hooks.
5.1.3	Each hook should be able to support at least 2 kg
5.1.4	A star base with at least five (5) anti-static castors wheels of at least 50mm diameter.
5.1.5	At least 2 castors should come equipped with brakes.
5.1.6	Height adjustable at least between 1400 and 2000mm.
5.1.7	Weighted base with at least 4 kg.
5.2	<u>Support Services</u>
5.2.1	Manufacturer's warranty: a) Duration 24 months from final receipt covering all manufacturing defects. b) The warranty covers all delivered equipment and the accessories delivered.

No	UNOPS minimum technical requirements
5.2.2	<p>Warranty with scope extended by the bidder:</p> <ul style="list-style-type: none"> a) Duration 24 months from final receipt covering all manufacturing defects. b) It includes corrective maintenance: labour and spare parts with no call limitations with the sole exclusion of damage due to proven misuse. c) The corrective maintenance technical service must be performed by a company based in the country of destination. d) Corrective maintenance during warranty can be carried out at the beneficiary premise, or by mailing the broken equipment at supplier or manufacturer premise and mailing back the repaired equipment or the substituting one to the Beneficiary premise. In the case of transportation all the costs and risks will be covered by the Supplier. The time elapsed between the communication of the faulty equipment and the intervention on site will be, within the warranty period, no more than 2 business days. e) The Supplier will notify the end user in case of any equipment Recall / Field safety Notice / Alerts as issued by the manufacturer/NRA or any other relevant authorities and organizations
5.3	<u>Documents to be presented with the offer</u>
5.3.1	Bidder declaration of the brand name, manufacturer, Country of Origin and model name of all the goods (including all consumables and accessories included in the offer)
5.3.2	Manufacturer's ISO 13485 certificate or ISO 9001 when ISO13485 is not applicable.
5.3.3	<p>Market authorization for one of the following markets/requested certifications:</p> <ul style="list-style-type: none"> i. EUROPE/CE 2017/745 conformity or equipment with CE mark according with 93/42 EEC ii. USA/FDA certification (Food and Drug Administration) for internal US market iii. CANADA/SOR/98-282 conformity certification iv. AUSTRALIA/TGA Conformity certification v. JAPAN/ PMDA pre-market approval
5.3.4	Certificate of registration/Authorization/ Waiver the Import and use of the device by the health regulatory authority (NRA) of the Country of destination.
5.3.5	Factory certificate that the equipment, specific offered model, is not discontinued on the production line.
5.3.6	Manufacturer's Authorization: Bidders may submit a simple copy of the manufacturer's authorization within their bid, indicating the offered model, according to the bidder's own form.
5.3.7	Manufacturer's technical marketing catalogues. Manufacturer marketing catalogues of the goods offered to verify compliance with the technical specifications. Catalogues must be provided in English.
5.3.8	Bidder shall provide invoice/proforma invoice, packing list and at least a draft shipping document without indicating the word "Draft" and with fixed BL/AWB number to UNOPS at least one month in advance before the actual shipping date for the application of Tax Exemption Certificate (TEC).
5.4	<u>Payment Terms</u>
5.4.1	<ul style="list-style-type: none"> • 80% of the cost of the goods shall be paid upon delivery per the Incoterms 2020 • 20% of the cost of the goods and 100% of the cost of associated services shall be paid upon final acceptance of the goods and the associated services (delivery, installation, testing, commissioning, training of users on operation), as applicable.
5.4.2	Sample Photo - The images below are intended solely as a guiding support and should be considered as purely indicative and not restrictive of the expected item characteristics:


No	UNOPS minimum technical requirements
	

Delivery requirements for IV Stand:

Delivery time	Bidder shall deliver the goods within 10 weeks after issuance of Purchase Order to UNOPS warehouse in Yangon, Myanmar (excluding the time for custom clearance process).
Delivery place and Incoterms rules	UNOPS warehouse in Yangon, Myanmar. Delivered at Place Unloaded [DPU] Incoterms 2020
Consignee details	UNOPS Myanmar No.12/O , Pyi Thu Lane, 7 Miles, Mayangone T/S, Yangon, Myanmar
UNOPS Right to vary requirements	At the time the Contract is awarded, UNOPS reserves the right to vary the quantity of the goods and associated services specified above, provided this does not exceed +/- 20%, without any change in the unit prices or other terms and conditions of the ITB.
Pre-inspection	The goods will be pre-inspected by UNOPS or by a third party appointed by UNOPS either in the production site or in Yangon.

Lot No 6: Cardiac table - Quantity: 346 EA
Technical specifications for Goods:

No	UNOPS minimum technical requirements
6	<u>Cardiac table</u> Board used for cardiopulmonary resuscitation (CPR) compressions.
6.1	<u>Technical requirements</u>
6.1.1	Rigid board.
6.1.2	Solid surface.
6.1.3	At least three (3) handle holes.
6.1.4	Made of ABS, acrylic, PP or better plastic resistant material.
6.1.5	Lightweight no greater than 2.5 kg
6.1.6	Anatomically designed.
6.1.7	Built-in handles
6.1.8	Possibly to be mounted in the emergency cart.
6.1.9	Load capacity not less than 160 kg.
6.2	<u>Dimensions:</u>
6.2.1	LxWxH: 60 x 45 x 2 cm approximately (+/-5%).
6.3	<u>Support Services</u>
6.3.2	<u>Manufacturer's warranty:</u>

No	UNOPS minimum technical requirements
	<p>a. Duration of 24 months from final receipt covering all manufacturing defects.</p> <p>b. The warranty covers all delivered equipment and the accessories delivered.</p> <p>c. It includes corrective maintenance: labour and spare parts with no call limitations with the sole exclusion of damages due to proven misuse.</p> <p>d. Corrective maintenance during warranty can be carried out at the beneficiary premise, or by mailing the broken equipment at supplier or manufacturer premise and mailing back the repaired equipment or the substituting one to the Beneficiary premise. In the case of transportation all the costs and risks will be covered by the Supplier. The time elapsed between the communication of the faulty equipment and the intervention on site will be, within the warranty period, no more than 2 business days.</p>
6.4	<u>Documents to be presented with the offer</u>
6.4.1	Bidder declaration of the brand name, manufacturer, Country of Origin and model name of all the goods (including all consumables and accessories included in the offer).
6.4.2	Manufacturer's ISO 13485 certificate or ISO 9001 when ISO13485 is not applicable.
6.4.3	Market authorization for one of the following markets/requested certifications: <p>a. EUROPE/CE 2017/745 conformity or equipment with CE mark according with 93/42 EEC.</p> <p>b. USA/FDA certification (Food and Drug Administration) for internal US market.</p> <p>c. CANADA/SOR/98-282 conformity certification.</p> <p>d. AUSTRALIA/TGA Conformity certification.</p> <p>e. JAPAN/ PMDA pre-market approval.</p>
6.4.4	Certificate of registration/Authorization/ Waiver the Import and use of the device by the health regulatory authority (NRA) of the Country of destination.
6.4.5	Manufacturer's Authorization: Bidders may submit a simple copy of the manufacturer's authorization within their bid, indicating the offered model, according to the bidder's own form.
6.4.6	Manufacturer's technical marketing catalogues. Manufacturer marketing catalogues of the goods offered to verify compliance with the technical specifications. Catalogues must be provided in English.
6.4.7	Bidder shall provide invoice/proforma invoice, packing list and at least a draft shipping document without indicating the word "Draft" and with fixed BL/AWB number to UNOPS at least one month in advance before the actual shipping date for the application of Tax Exemption Certificate (TEC).
6.5	<u>Payment Terms</u>
6.5.1	<ul style="list-style-type: none"> 80% of the cost of the goods shall be paid upon delivery per the Incoterms 2020 20% of the cost of the goods and 100% of the cost of associated services shall be paid upon final acceptance of the goods and the associated services (delivery, installation, testing, commissioning, training of users on operation), as applicable.
6.5.2	Sample Photo - The images below are intended solely as a guiding support and should be considered as purely indicative and not restrictive of the expected item characteristicsExample:
	

Delivery requirements for Cardiac Table:


Delivery time	Bidder shall deliver the goods within 10 weeks after issuance of Purchase Order to UNOPS warehouse in Yangon, Myanmar (excluding the time for custom clearance process).
----------------------	--

Delivery place and Incoterms rules	UNOPS warehouse in Yangon, Myanmar. Delivered at Place Unloaded [DPU] Incoterms 2020
Consignee details	UNOPS Myanmar No.12/O , Pyi Thu Lane, 7 Miles, Mayangone T/S, Yangon, Myanmar
UNOPS Right to vary requirements	At the time the Contract is awarded, UNOPS reserves the right to vary the quantity of the goods and associated services specified above, provided this does not exceed +/- 20%, without any change in the unit prices or other terms and conditions of the ITB.
Pre-inspection	The goods will be pre-inspected by UNOPS or by a third party appointed by UNOPS either in the production site or in Yangon.

Lot No 7: Wheelchair, adult - Quantity: 1348 EA

Technical specifications for Goods:

No	UNOPS minimum technical requirements
7	<u>Wheelchair, adult</u> Mechanical wheelchair for adult patients, used for patient transportation purposes in hospital wards.
7.1	<u>Operational characteristics</u>
7.1.1	Chrome plated strong tubular steel frame.
7.1.2	Fixed padded armrests.
7.1.3	Vinyl or similarly upholstery on seat, arms and backrest.
7.1.4	Vinyl-upholstered seat and back: antibacterial, antistatic, flame-retardant, resistant to corrosion of water, detergent soap, ethyl alcohol solution and to hypochlorite of sodium.
7.1.5	Swinging, detachable footrests with nylon heel loops.
7.1.6	Two spaced rear wheels with circular handles for patient self-propulsion.
7.1.7	Two small ball bearing swivel castors in the front.
7.1.8	Non-marking rubber tires.
7.1.9	Wheel locks for rear wheels.
7.1.10	Capacity at least up to 150 kg.
7.1.11	Two handles with hand-grip at top rear backrest.
7.1.12	Seat not less than 460 x 400 mm (WxD)
7.1.13	Overall dimensions not exceeding 950x700x1350 mm (HxWxL) with footrest fully extended.
7.1.14	Preferable foldable design for ease of storage and transportation.
7.2	<u>Accessories to be included in the offered price:</u>
7.2.1	An extra pair (2 pcs) of tubes in case tires (6.1.8) have tubes and an adapter allowing to inflate them with any pump.
7.3	<u>Support Services</u>
7.3.1	Equipment installation: the equipment must be delivered at destination, already assembled and tested.
7.3.2	Training on use and maintenance: Operating, cleaning, daily maintenance and training instructions should be included in the manuals delivered with the equipment.
7.3.3	Manufacturer's warranty: a) Duration 24 months from final receipt covering all manufacturing defects. b) The warranty covers all delivered equipment and the accessories delivered.
7.3.4	Warranty with scope extended by the bidder: a) Duration 24 months from final receipt covering all manufacturing defects. b) It includes corrective maintenance: labour and spare parts with no call limitations with the sole exclusion of damage due to proven misuse. c) The preventive and corrective maintenance technical service must be performed by a company based in the country of destination. d) Corrective maintenance during warranty can be carried out at the beneficiary premise, or by mailing the broken equipment at supplier or manufacturer premise and mailing back the

	<p>repaired equipment or the substituting one to the Beneficiary premise. In the case of transportation all the costs and risks will be covered by the Supplier. The time elapsed between the communication of the faulty equipment and the intervention on site will be, within the warranty period, no more than 2 business days.</p> <p>e) The Supplier will notify the end user in case of any equipment Recall / Field safety Notice / Alerts as issued by the manufacturer/NRA or any other relevant authorities and organizations</p>
7.4	<u>Documents to be presented with the offer</u>
7.4.1	Bidder declaration of the brand name, manufacturer, Country of Origin and model name of all the goods (including all consumables and accessories included in the offer)
7.4.2	Manufacturer's ISO 13485 certificate or ISO 9001 when ISO13485 is not applicable.
7.4.3	<p>Market authorization for one of the following markets/requested certifications:</p> <ul style="list-style-type: none"> a) EUROPE/CE 2017/745 conformity or equipment with CE mark according with 93/42 EEC b) USA/FDA certification (Food and Drug Administration) for internal US market c) CANADA/SOR/98-282 conformity certification d) AUSTRALIA/TGA Conformity certification e) JAPAN/ PMDA pre-market approval
7.4.4	Certificate of registration/Authorization/ Waiver the Import and use of the device by the health regulatory authority (NRA) of the Country of destination.
7.4.5	Manufacturer's Authorization: Bidders may submit a simple copy of the manufacturer's authorization within their bid, indicating the offered model, according to the bidder's own form.
7.4.6	Manuals: Copies of User and Maintenance manual must be submitted in digital format in English.
7.4.7	Manufacturer's technical marketing catalogues. Manufacturer marketing catalogues of the goods offered to verify compliance with the technical specifications. Catalogues must be provided in English.
7.4.8	Letter of the manufacturer stating his commitment of spare parts availability for 5 years in case of award (only availability the supply of spare parts shall not be included in the bid price).
7.4.9	Bidder shall provide invoice/proforma invoice, packing list and at least a draft shipping document without indicating the word "Draft" and with fixed BL/AWB number to UNOPS at least one month in advance before the actual shipping date for the application of Tax Exemption Certificate (TEC).
7.5	<u>Payment Terms</u>
7.5.1	<ul style="list-style-type: none"> • 80% of the cost of the goods shall be paid upon delivery per the Incoterms 2020 • 20% of the cost of the goods and 100% of the cost of associated services shall be paid upon final acceptance of the goods and the associated services (delivery, installation, testing, commissioning, training of users on operation), as applicable.
7.5.2	Sample Photo - The images below are intended solely as a guiding support and should be considered as purely indicative and not restrictive of the expected item characteristics:
	

Delivery requirements for wheelchair:

UNOPS Requirements	
Delivery time	Bidder shall deliver the goods within 10 weeks after issuance of Purchase Order to UNOPS warehouse in Yangon, Myanmar (excluding the time for custom clearance process).
Delivery place and Incoterms rules	UNOPS warehouse in Yangon, Myanmar. Delivered at Place Unloaded [DPU] Incoterms 2020
Consignee details	UNOPS Myanmar No.12/O , Pyi Thu Lane, 7 Miles, Mayangone T/S, Yangon, Myanmar
UNOPS Right to vary requirements	At the time the Contract is awarded, UNOPS reserves the right to vary the quantity of the goods and associated services specified above, provided this does not exceed +/- 20%, without any change in the unit prices or other terms and conditions of the ITB.
Pre-inspection	The goods will be pre-inspected by UNOPS or by a third party appointed by UNOPS either in the production site or in Yangon.

Related services requirements for wheelchairs

SN	UNOPS minimum requirements for services	Place where services will be performed	Q-ty	Physical unit
1	Assemble at destinations	Yangon General Hospital	133	EA
2	Assemble at destinations	Mandalay General Hospital	100	EA
3	Assemble at destinations	North Okkala General Hospital	53	EA
4	Assemble at destinations	Naypyitaw General Hospital	67	EA
5	Assemble at destinations	Yangon Specialist Hospital (500-bed)	33	EA
6	Assemble at destinations	Thingangyun General Hospital	33	EA
7	Assemble at destinations	Magway General Hospital, Magway Region	13	EA
8	Assemble at destinations	Insein General Hospital	33	EA
9	Assemble at destinations	Magway Teaching Hospital	13	EA
10	Assemble at destinations	Sao San Tun General Hospital	33	EA
11	Assemble at destinations	Women and Children Hospital, Shan South State	13	EA
12	Assemble at destinations	300-bed Mandalay Teaching Hospital	20	EA
13	Assemble at destinations	Myitkyina General Hospital, Kachin State	33	EA
14	Assemble at destinations	Bamaw General Hospital, Kachin State	13	EA
15	Assemble at destinations	Loikaw General Hospital, Kayah State	33	EA
16	Assemble at destinations	Hpa-an General Hospital, Kayin State	13	EA
17	Assemble at destinations	Hakha General Hospital, Chin State	13	EA
18	Assemble at destinations	Monywa General Hospital, Sagaing Region	33	EA
19	Assemble at destinations	Kalay General Hospital, Sagaing Region	20	EA
20	Assemble at destinations	Sagaing General Hospital, Sagaing Region	13	EA
21	Assemble at destinations	Dawei General Hospital, Tanintharyi Region	13	EA
22	Assemble at destinations	Myeik General Hospital, Tanintharyi Region	13	EA

SN	UNOPS minimum requirements for services	Place where services will be performed	Q-ty	Physical unit
23	Assemble at destinations	Bago General Hospital, Bago Region	33	EA
24	Assemble at destinations	Pyay General Hospital, Bago Region	33	EA
25	Assemble at destinations	Taunggu General Hospital, Bago Region	13	EA
26	Assemble at destinations	Minbu General Hospital, Magway Region	13	EA
27	Assemble at destinations	Pakkoku General Hospital, Magway Region	13	EA
28	Assemble at destinations	Pyinoolwin General Hospital, Mandalay Region	20	EA
29	Assemble at destinations	Meikthtila General Hospital, Mandalay Region	13	EA
30	Assemble at destinations	Mawlamyine General Hospital, Mon State	33	EA
31	Assemble at destinations	Mawlamyine Women and Children Hospital, Mon State	13	EA
32	Assemble at destinations	Sittwe General Hospital, Rakhine State	33	EA
33	Assemble at destinations	East Yangon General Hospital, Yangon Region	13	EA
34	Assemble at destinations	West Yangon General Hospital, Yangon Region	13	EA
35	Assemble at destinations	Hlaingtharyar General Hospital, Yangon Region	13	EA
36	Assemble at destinations	Thanlyin General Hospital, Yangon Region	13	EA
37	Assemble at destinations	Lashio General Hospital, Shan North State	33	EA
38	Assemble at destinations	Kengtong General Hospital, Shan East State	13	EA
39	Assemble at destinations	Patheingyi General Hospital, Ayeyarwady Region	33	EA
40	Assemble at destinations	Pha-Lan General Hospital, Chin State	13	EA
41	Assemble at destinations	Shwe Bo General Hospital, Sagaing Region	13	EA
42	Assemble at destinations	Kyaukse General Hospital, Mandalay Region	13	EA
43	Assemble at destinations	Nyaung-U General Hospital, Mandalay Region	13	EA
44	Assemble at destinations	Kyaukpadaung General Hospital, Rakhine State	13	EA
45	Assemble at destinations	Loilem General Hospital, Shan South State	13	EA
46	Assemble at destinations	Maubin General Hospital, Ayeyarwady Region	13	EA
47	Assemble at destinations	Myaungmya General Hospital, Ayeyarwady Region	13	EA
48	Assemble at destinations	Laputta General Hospital, Ayeyarwady Region	13	EA
49	Assemble at destinations	Pyapon General Hospital, Ayeyarwady Region	13	EA
50	Assemble at destinations	Hinthada General Hospital, Ayeyarwady Region	13	EA
51	Assemble at destinations	Pyinmana General Hospital, Naypyitaw Council	13	EA
52	Assemble at destinations	Naypyitaw General Hospital (300-bed)	20	EA
53	Assemble at destinations	Tedim Township Hospital	7	EA
54	Assemble at destinations	Mindat District Hospital	7	EA
55	Assemble at destinations	Mogoke Township Hospital	7	EA
56	Assemble at destinations	Yamethin District Hospital	7	EA

SN	UNOPS minimum requirements for services	Place where services will be performed	Q-ty	Physical unit
57	Assemble at destinations	Kawthaung District Hospital	7	EA
58	Assemble at destinations	Ye Township Hospital	7	EA
59	Assemble at destinations	Thaton District Hospital	7	EA
60	Assemble at destinations	Thandwe District Hospital	7	EA
61	Assemble at destinations	Tachileik District Hospital	7	EA
62	Assemble at destinations	Ann District Hospital	5	EA
63	Assemble at destinations	Moe Nyin District Hospital	5	EA
64	Assemble at destinations	Kutkai General Hospital -Shan North	5	EA
65	Assemble at destinations	Shadaw General Hospital - Kayah State	5	EA
Total quantity			1,348	EA