

Amendment No.1
ITB YEM-007/2012
Procurement of PCR Accessories

This Amendment is raised to:

- 1- Extend the deadline to read Monday 2 July 2012 10:00 am Sana'a local time instead of 30 June 2012.
- 2- The public bid opening will be on Monday 2 July 2012 at 11:00 am
- 3- Answer the following clarification received from suppliers.

Question 1:

Item 1: Bench centrifuge: kindly confirm that you are asking for a rotor for 0.5 ml instead of 0.5 µl and adapter for 0.2 ml instead of 0.2 µl.

Answer 1:

The correct specs shall read as follows:

Microcentrifuge (0 to 15,000rpm) supplied with fixed angle rotor for 1.5 ml tubes and adaptors for 0.5ml and 0.2ml (220V 50/60Hz)

Question 2:

For items 9 to 14: please confirm that these micropipettes are fixed volume and not variable volume.

Answer 2:

The volumes are variable as shown in the below table:

Item No.	Item
9	variable micropipette, 100 to 1000 µl
10	variable micropipette, 20 to 200 µl
11	variable micropipette, 10 to 100 µl
12	variable micropipette, 1 to 20 µl
11	variable micropipette, 0.5 to 10 µl
14	variable micropipette, 0.0 to 5 µl

Question No.3

Is it possible to send the offer by email?

Answer 3:

Offers received by email or fax will be rejected

Handwritten signature and date:
27/6/2012