

ANNEX-B-c
SPECIFICATION OF MEDICAL EQUIPMENT
DELIVERY LOCATION AT TRAUMA CENTRE AT QUETTA.

DIATHERMY MACHINE WITH SPECIFICATIONS

No	Item	Specifications	Details	QTY
1.		Output Power	500 Watts	2
		Power Selection	5 Steps	
		Cooling Fan		800
		Timer	0—30min	
		Frequency	27.12 MHZ	
		Wave Length	11mm	
		Input Supply	230 V Ac.50Hz	
		Size LxWxH	16"x12"x29"	
		Net Weight	20Kg	

ELECTRICAL OPERATING EXAMINATION COUCH WITH SPECIFICATIONS

No	Item	Specifications	Details	QTY
1.		Model	HS5236	4
		Functions	Multi-Function	
		Foldable	Foldable	
		Type	Examination	
		Material	Synthetic Leather	
		Size LxWxH	1800x690x970 mm	
		Made /Origin	China	

Operation Theatre Equipment

S/N	Particular	Department	Specs	QTY
1.	Laryngoscopes-Video	Anaesthesia	3.0"screen LCD display.	2
			Screen resolution 960*480	
			Compatible with all POSTMAN series laryngoscope.	
			Tilt up and down 135°.	
			Support SD Card storage, max up to 128G capacity.	
			Support photo and video recording function.	
			Camera Resolution:1080P.	
			With rechargeable battery	

No The Biomedical Equipments shall met the Certification (FDA/CE/MHLW) where applicable
Country of Manufacture: Preferably - USA/Europe/Japan