

Section II: Schedule of Requirements

eSourcing reference: ITB/2023/49570

Technical specifications for Goods and Comparative Data Table:

Lot 01:

Item 01: Supply & delivery of 1,600 Nos of X-Ray Film Size 14" x 17"

Item No	Minimum Requirement
1	Manufacturer Name
2	Model No
3	Country of Origin
4	Operational Requirements
4.1	Size 14" x 17"
4.2	Medical Dry Laser Imaging Film,
4.3	Maximum Density: up to 3.6
4.4	Light-sensitive protective Layer
4.5	Use: Specially in CR, CT, MRI, DSA, Mammography, OPG & others Medical Imaging modalities
5	Warranty and maintenance
5.1	Warranty period of minimum 12 months should be given starting from the date of delivery
6	Documentation
6.1	User, technical and maintenance manuals must be supplied in English language
7	Quality Requirements
7.1	If required , the Supplier has to arrange the Marketing Authorization/NOC/Registration from the Directorate General of Drug Administration (DGDA) Bangladesh after the issuance of the PO.
7.2	Have regulatory approval and marketing authorization issued by one of the GHTF Founding Member countries (the European Union, the United States, Canada, Japan and Australia) or be prequalified by the WHO (Ex; CE, FDA...)
7.3	Manufacturer's QMS ISO13485 or ISO9001, QMS shall be issued by CABs, Notified or Accredited bodies recognized by the Regulatory Authority of one of the GHTF Founding Member countries (the European Union, the United States, Canada, Japan and Australia) and shall be recognized by such Authorities.
8	Delivery and storage

8.1	Delivery within 4 weeks of the issuance of the PO
8.2	Delivery to NTP Central Warehouse, 250 Bedded TB Hospital, Shyamoli, Dhaka
8.3	Incoterm: DAP
9	Contact details of Local service agent should be submitted
9.1	<i>Name</i>
9.2	<i>Address</i>
9.3	<i>Contact number</i>

Item 02: Supply & delivery of 28,350 Nos of X-Ray Film Size 10" x 14 "

Item No	Minimum Requirement
1	Manufacturer Name
2	Model No
3	Country of Origin
4	Operational Requirements
4.1	Size 10" x 14 "
4.2	Medical Dry Laser Imaging Film,
4.3	Maximum Density: up to 3.6
4.4	Light-sensitive protective Layer
4.5	Use: Specially in CR, CT, MRI, DSA, Mammography, OPG & others Medical Imaging modalities
5	Warranty and maintenance
5.1	Warranty period of minimum 12 months should be given starting from the date of delivery
6	Documentation
6.1	User, technical and maintenance manuals must be supplied in English language
7	Quality Requirements
7.1	If required , the Supplier has to arrange the Marketing Authorization/NOC/Registration from the Directorate General of Drug Administration (DGDA) Bangladesh after the issuance of the PO.
7.2	Have regulatory approval and marketing authorization issued by one of the GHTF Founding Member countries (the European Union, the United States, Canada, Japan and Australia) or be prequalified by the WHO (Ex; CE, FDA...)
7.3	Manufacturer's QMS ISO13485 or ISO9001, QMS shall be issued by CABs, Notified or Accredited bodies recognized by the Regulatory Authority of one of the GHTF Founding Member countries (the

	European Union, the United States, Canada, Japan and Australia) and shall be recognized by such Authorities.
8	Delivery and storage
8.1	Delivery within 4 weeks of the issuance of the PO
8.2	Delivery to NTP Central Warehouse, 250 Bedded TB Hospital, Shyamoli, Dhaka
8.3	Incoterm: DAP
9	Contact details of Local service agent should be submitted
9.1	<i>Name</i>
9.2	<i>Address</i>
9.3	<i>Contact number</i>

Item 03: Supply & delivery of 1,550 Nos of X-Ray Film Size 8" x 10"

Item No	Minimum Requirement
1	Manufacturer Name
2	Model No
3	Country of Origin
4	Operational Requirements
4.1	Size 8" x 10"
4.2	Medical Dry Laser Imaging Film,
4.3	Maximum Density: up to 3.6
4.4	Light-sensitive protective Layer
4.5	Use: Specially in CR, CT, MRI, DSA, Mammography, OPG & others Medical Imaging modalities
5	Warranty and maintenance
5.1	Warranty period of minimum 12 months should be given starting from the date of delivery
6	Documentation
6.1	User, technical and maintenance manuals must be supplied in English language
7	Quality Requirements
7.1	If required , the Supplier has to arrange the Marketing Authorization/NOC/Registration from the Directorate General of Drug Administration (DGDA) Bangladesh after the issuance of the PO.

7.2	Have regulatory approval and marketing authorization issued by one of the GHTF Founding Member countries (the European Union, the United States, Canada, Japan and Australia) or be prequalified by the WHO (Ex; CE, FDA...)
7.3	Manufacturer's QMS ISO13485 or ISO9001, QMS shall be issued by CABs, Notified or Accredited bodies recognized by the Regulatory Authority of one of the GHTF Founding Member countries (the European Union, the United States, Canada, Japan and Australia) and shall be recognized by such Authorities.
8	Delivery and storage
8.1	Delivery within 4 weeks of the issuance of the PO
8.2	Delivery to NTP Central Warehouse, 250 Bedded TB Hospital, Shyamoli, Dhaka
8.3	Incoterm: DAP
9	Contact details of Local service agent should be submitted
9.1	<i>Name</i>
9.2	<i>Address</i>
9.3	<i>Contact number</i>

Delivery requirements and Comparative Data Table:

UNOPS Requirements	
Delivery time	Bidder shall deliver the goods (all 3 items) within 04 weeks after Purchase Order / Contract signature.
Delivery place and Incoterms rules	<p>Delivery place:</p> <p>NTP Central Warehouse, 250 Bedded TB Hospital, Shyamoli, Dhaka</p> <p>Incoterm: DAP</p> <p>(Director General of Health Services of the Ministry of Health and Family Welfare, Government of Bangladesh will facilitate custom clearance support including waivers for VAT, Duty and Custom Clearance charges)</p>
Consignee details	<p>D.G. Central Medical Stores Dept</p> <p>30, Shahid Tajuddin Ahmed Sarani, Tejgaon, Dhaka 1208, Bangladesh</p> <p>BIN: 002702482-0203</p>