

03 July 2024

Requests for Clarifications No. 5
UNFPA/DNK/ITB/24/006

1. Question 1

In relation to Item 9. Vacuum Extractor

"With foot pedal and manual suction function activation. Pedal provided". Is the foot pedal required as an on/ off for the vacuum or is a foot vacuum regulator required?

UNFPA Response: An on/off type pedal is acceptable, but the vacuum extractor must have an adjustable vacuum regulator.

2. Question 2

Regarding Expendable items for 2 years:

Could you advise if we can quote a unite price for each expendable item? if not could you please provide us with the estimated quantities of expendable items for 2 years?

UNFPA Response: Please refer to **UNFPA Response to Question 15** in **Requests for Clarifications No. 2** document as published on UNGM: Bidders should be guided by the amounts requested in the ITB and quote for each expendable item in the separate line in the Price schedule, so UNFPA is able to estimate total cost of the ownership in the first 2 years. Please note that consumables can be shipped separately later, however for bidders it is very important to keep in mind freight charges and necessity to quote FCA and CPT prices.

3. Question 3

Regarding Annex II Price schedule form you have mentioned for

TOTAL FIRM FCA PRICE

TOTAL FIRM CPT PRICE

TOTAL PRICE FOR SERVICES (if applicable)

Could please explain if the total firm FCA price includes the total price for the device and total price for all expected expendables quantities? The same question for total firm CPT price.

UNFPA Response: The total firm FCA/CPT price must include costs for both medical devices and expendables. It is not possible to submit Offers for medical devices only without related expendables. Bidders must take into account the cost of split shipments, if any.