

RFQ628139 -2-RO Long term agreement for Digital Radiographic Mobile X-ray System

Clarifications for bidders No.1

	CLARIFICATION REQUEST	ANSWER
1	Could you please specify what changes were implemented in the re-launched RFQ in comparison to the initial one?	<p>All changes have been listed in the cover letter attached to this RFQ. Please find below a summary of the changes highlighted in red:</p> <ol style="list-style-type: none"> 1. 2.4.4. Selection Criteria in the special instruction for bidders were amended: <i>TOTAL cost over 7 years for 105 Xray Systems including costs of:</i> <ul style="list-style-type: none"> - Equipment + - <i>ONLINE training +</i> - <i>Onsite maintenance services for 1 year as part of the warranty period +</i> - Cost of rotating stock over 7 years of minimum 3 Xray systems + - Other costs (as deemed mandatory by the bidder) 2. The commercial offer template was amended to reflect the above changes. 3. 2.4.1. Qualification requirements in the special instruction for bidders were amended: <ul style="list-style-type: none"> - Operation: The Bidder shall have at <i>least five (5) years of operation</i> in the relevant technical field (radiology and medical imaging) at the international level. - References: <i>A minimum of five (5) satisfactory verifiable references</i> evidencing successful delivery of similar equipment within the past five (5) years. - Delivery time proposed: Bidder to provide lead-time in weeks for an FCA supplier's warehouse readiness (maximum acceptable lead-time FCA supplier's warehouse readiness is 6 weeks). - Rotating stock: Provision of a rotating stock of minimum 3 mobile X-rays system available for immediate dispatch (confirm availability of rotating stock of minimum 3 mobile X-rays systems, confirm pricing was provided in commercial offer, confirm location where the rotating stock is stored) 4. Qualification compliance matrix has been added, to be filled in by bidders. 5. Requirement of maintenance service 12.5 was amended.

		<p><i>The Contractor shall ensure that a suitable qualified person shall be available online within forty-eight (48) hours and within 2 weeks onsite following an unexpected breakdown and shall investigate the problem within the next twenty-four (24) hours during the warranty period.</i></p> <p>6. Minor words correction for requirements 4.1 and 8.2 in the technical specifications</p> <p>7. Technical requirements 4.3.1 and 4.3.2 were updated based on the clarifications received on the first round of the RFQ.</p>
--	--	--