

**RFQ628139 -2-RO Long term agreement for Digital Radiographic Mobile X-ray System**

**Clarifications for bidders No.3**

	<b>CLARIFICATION REQUEST</b>	<b>ANSWER</b>
1	Regarding the system packing (point 6 of “technical specification” document), we would like a confirmation if the following values for “storage” condition are accepted: ambient temperature of -10 to +55 °C and relative humidity (RH) of 20–80%.	These conditions could be accepted only if the supplier provides a certificate stating that these temperature conditions won’t affect the appropriate function of the X-ray equipment.