

## Clarification Document UNFPA/DNK/RFP/24/011

<b>BIDDER QUESTION</b>	<b>UNFPA RESPONSE</b>
Could you utilize instead existing LTAs that you have with our company for this procurement?	The current RFP has been set as an international open tender, thus we cant unfortunately utilize any existing LTAs, and would encourage you to submit your bid to <a href="mailto:bidtender@unfpa.org">bidtender@unfpa.org</a> before the deadline
Due to the complexity of the tender, we would highly appreciate if you could extend the tender deadline for additional two weeks, so we would have sufficient time to work on the proposals.	Due to the urgency of the tender we have intended to keep the tendering period to a minimum, whilst allowing bidders sufficient time to prepare their proposals. We have however after internal consultations with all relevant parties within UNFPA decided to extend the submission deadline until <b>August 7<sup>th</sup>, 10:00 Copenhagen time. We have also extended the deadline for asking questions to 31-07-24, at 16:00 Copenhagen time.</b>
It will be difficult to fit all the equipment on the list into the mobile clinic. Is it an option to extend the total length of the clinic from what is specified in the tender documents?	The tender document provides an indicative vehicle length of 10-12m long to accomodate a box body of approximately 8m long. The use of the words indicative and approximate are intended to convey flexibility of design and are appropriate for a request for proposal rather than a more rigid bid requirement. Proposals should not loose sight of the need to retain an appropriate level of mobility in an area that can have challenging conditions.
There are many pieces of equipment on the list which is not allocated to a specific room. Can we please have this information to finalize the design and provide the best possible solution?	The RFP describes functional 'spaces' rather than rooms and provides for hardwall, folding and curtain type partitions in order to allow maximum flexibility. Within this broad description of function and resources we would recommend that proposals include a logical distribution of equipment according to their design. This can be clarified during the evaluation process.
Can you provide more insight in the preferred delivery time. And pls provide clarity if delivery time is taken into account on the scoring.	Please see the amended invitation letter and submission forms, the required delivery lead time has been added: 4-6 weeks after UNFPA PO submission.