

QUESTIONS AND ANSWERS

BID: UNFPA/DNK/ITB/24/014

No: 01

Date: 2024.8.30

№	Question	Answer
1	<p>You mentioned: “Regulatory approvals required: Valid Mongolian official Import license from the main manufacturer for receiving, installing, and maintaining medical equipment in Mongolia” in the technical specs in the tender.</p> <p>Could you please explain your requested document?</p> <p>Are you asking registration document of the device in the Ministry of Health of Mongolia? Or</p> <p>Are you asking an authorization letter from the manufacturer?</p>	<p>The supplier should submit the following documents together with their bids:</p> <ol style="list-style-type: none"> 1. Special license of the Ministry of Health of Mongolia authorizing companies for importing, supplying and selling medical items within the country. It should be submitted along with the bid. 2. Manufacturer’s authorization letter to sell the product for this tender. It should be submitted together with the bidding document. <p>Please refer to: UNFPA/DNK/ITB.24.014 – 10.2.c <i>In the case of a Bidder not doing business within the country of destination, the Bidder is or will be represented by an Agent in the country that is equipped and able to carry out the supplier’s maintenance, training, repair and spare parts-stocking obligations prescribed in the Section II, Technical Specifications and Schedule of Requirements</i></p>
2	<p>Could you please clarify below technical specifications? Item14/OR22 Multiparametric TRANSPORT patient monitor for adult and pediatric patients – Is its English language or Russia? (Indications and messages on the equipment must be in Russian language.)</p>	<p>Indications and messages on the equipment must include English language.</p>
3	<p>Item21/OR31 “Patient ventilator &Respirator” – Is it Anesthesia machine or Ventilator/respirator? Technical specification is showing some anesthesia information (Device intended for use as an anesthesiology workstation) and Accessories is included Two (2) vaporizers, one for Sevoflurane and one for Isoflurane. With a visual indicator of the filling level of the anesthetic agent, and adapter for filling the vaporizer if required.</p>	<p>The device is a Ventilator machine specifically tailored for anesthesiology procedures (Adult & pediatric). Anesthesia Machines comply the item requirements.</p>
4	<p>We are keen to participate in the subject tender and seek clarification regarding the regulatory approvals required. Specifically, we would like to confirm if a valid Mongolian official import license from the main manufacturer for receiving, installing, and maintaining medical equipment in Mongolia is a mandatory requirement.</p> <p>Additionally, if this is indeed mandatory,</p>	<p><i>Please refer to UNFPA’s answer for Question #1</i></p>

	could you please clarify whether the document needs to be submitted at the time of bid submission, or if it is acceptable to submit a declaration at the time of bid submission and provide the actual license after the contract has been awarded?	
5	We are reviewing the detailed technical specification requirement, and there is a requirement for valid Mongolian official Important License for each medical equipment. Sorry that we don't have this license, is it mandatory?	<i>Please refer to UNFPA's answer for Question #1</i>
6	As per our experience, the UNFPA should be in tax exemption in various countries.	Yes. UNFPA is exempted from tax in Mongolia. Therefore, the bid price shall be exclusive of direct taxes. UNFPA will issue a tax exemption letter to the supplier for their custom clearance process.