QUALITY CRITERIA

2024-08-12

RFQ 642487-MJD - REQ 148214 - Micro SPECT-CT for Preclinical Imaging for Mexico - MEX6013

Minimum Score Technically Acceptable:										
		Specification, SOW, or RFP Point Evaluated=>	**1) Proposed production, installation, testing and commissioning plan, (Bidder shall submit a plan outlining the timeline and/or milestones, checklist for these activities**	**2) Compliance with IAEA requirements and other mandatory requirements of the specification including detailed design based on Bidder's standard components (Bidder shall meet the mandatory requirements**	**3) Proposed training regime for staff, both end-users and first line maintenance staff**	**4) Evidence of good past performance for the successful implementation of micro-SPECT/CT systems (Bidder shall provide a list of at least 5 references from different countries/regions with similar micro-SPECT/CT systems provided in past 5 years)**	**5) Proposed spare parts and consumables package as applicable and 10 years availability of spare parts/consumables (Bidder shall submit a copy of service contract**	**6) Availability of technical support in the country/region and feasibility of early response on site (within 35 hrs), incl. support maintenance services, and a key account manager (Bidder shall submit a sample of Service Level Agreement and state the nearest authorized service center(s) available**	**7) Guarantee uptime of more than 93%**	**Weighted Score**
Bidder	**Reason for Rating / Technical Non-Acceptability**	**Weight of Area=> Evaluated (Must total 100%)**								