

Questions & Answers

INVITATION TO BID ref.# ITB/CAI/003/24

Provision of supply and delivery of MRI devices for the UNHCR-Egypt Operation

CLOSING DATE AND TIME FOR RECEIVING QUESTIONS:

15 October 2024 – 23:59 EGYPT TIME

No.	Questions	UNHCR Responses
1	<p>The specification below under the Gantry section needs to be amended.</p> <p>No 1.5T MRI gantry in the market weighs 200kg, it's requested to amend with "state magnet weight".</p> <p>Gantry weight is usually much more than 200KG (2000KG+) and the patient table max weight is 250KG in most of the vendors</p>	<p>It meant that the machine table capacity weight is 200Kgms.</p>
2	<p>Point. 1.11: by helium consumption?</p> <p>Is it meant how many liters of liquid helium does the machine have or the annual refilling consumption as now there is sealed magnets that doesn't need refilling?</p>	<p>It meant the least helium consumption</p>
3	<p>Point. 3.1 : Maximum amplitude, single axis, mT/m, at least true 44mT/m or higher per each axis (offers less than this figure will be rejected) ,</p> <p>Referring to the sentence of Maximum amplitude</p> <p>Gradient amplitude should be minimum not maximum as in the same line it mentioned at least 44 T/m</p>	<p>Minimum amplitude, single axis, mT/m, at least true 44mT/m or higher per each axis (offers less than this figure will be rejected)</p>
4	<p>Point. 3.2 : the slew rate of 200 mT/m/s is that meant to be true also like the amplitude???</p>	<p>Yes it meant True Slew rate</p>
5	<p>For the coils, is it accepted to offer a single coil that can cover more than one of the anatomies mentioned in the specs</p>	<p>Not accepted.</p> <p>Coils should be delivered as requested.</p>

END