

Clarification 1 – RFQ 642916 Dynamic Thorax Phantom (COS6027) #152566

4.2.1. Overall Dimensions (26" x 13" x 11" / 67 cm x 32 cm x 28 cm): Do you require the overall size of the phantom itself, the overall size of the phantom that will be scanned or the overall size of the phantom including the box?

IAEA: The dimensions and weight are only for reference.

4.2.2. Overall weight: 17.2 kg. : The overall weight of the phantom only or the phantom with its box?

IAEA: It's for the phantom without the box and it is acceptable to have differences of +/- 15%.

4.2.9. Maximum surrogate platform load: 5.4 kg : Do they mean Minimum instead of Maximum?

IAEA: The maximum surrogate platform load is also for reference. I can technically accept variations.