



United Nations Population Fund  
Next to the Lutheran Vocational  
School, Ar-Radwan Street, Beit  
Hanina, Jerusalem.  
Tel: 02 581 71 67 Ext. 71011  
Website: www.unfpa.ps

ITB 2016-001 Addendum No. 1 - dated 07 September 2016

**Addendum No.1**

**Reference: ITB No. UNFPA/PAL/16/001**

**Subject: Supply, Delivery, Installation and Training – Digital Mammography**

**Issue date: 22 August 2016**

**Closing date: 20 September 2016, at 10:00 a.m. (Jerusalem time).**

This correspondence pertaining to the above-mentioned project shall be considered as an integral part of the ITB/tender documents.

Below please find UNFPA's clarifications and an official reply to all inquiries raised by the participant bidders.

<b>QUESTION No. 1:</b>	What is the duration of the bid security?
<b>ANSWER No. 1:</b>	The bid security shall be valid for 90 days after the bid submission date.
<b>QUESTION No. 2:</b>	Image evaluation: Contrast enhancement (with table): What kind of table and for which purposes it is needed? Is this table for operator? Or for Patient?
<b>ANSWER No. 2:</b>	The Digital Mammography must include contrast enhancement feature for the image processing. We are not referring to a physical table.

Interested bidders shall acknowledge receipt of this addendum/pre-bid minutes by returning/including it, signed and stamped, with their bids.

For your kind attention and reference,

Mayyada Malki  
Operations Manager  
United Nations Population Fund

